



IntelliPlex™ Lung Cancer Panel – cfRNA

REF 82031 24 Reactions

RUO For Research Use Only



PlexBio Co., Ltd.
6F-1, No. 351, Yangguang St., Neihu Dist.
Taipei City, 11491, Taiwan
<http://www.plexbio.com>



IMPORTANT:
Read the instructions carefully prior to use

1. INTENDED USE

The **IntelliPlex Lung Cancer Panel - cfRNA**, based on π Code™ technology and PlexBio's instrument platform, is an in vitro molecular assay intended for the qualitative identification of 28 gene variants of the ALK, ROS1, RET, NTRK1, and MET genes using cell-free RNA derived from plasma of patients with non-small cell lung cancer (NSCLC). The kit is to be used by trained personnel in a professional laboratory environment. This product is for research use only, and not for use in diagnostic procedures.

2. INTRODUCTION

Studies of NSCLC has identified recurrent 'driver' mutations that occur in multiple oncogenes, including AKT1, ALK, BRAF, EGFR, HER2, KRAS, MEK1, MET, NRAS, PIK3CA, RET, and ROS1, and these markers serve as the basis for the molecular classification of NSCLC. Both DNA mutations and gene variants contribute to oncogenesis in NSCLC. Furthermore, copy number variations (CNVs) signatures have been developed to differentiate between NSCLC and SCLC, in particular for poorly differentiated NSCLC, to complement current pathology practices using small biopsies or cytology specimens. Targeted small molecule inhibitors are now available or being developed to benefit specific molecularly defined subsets of NSCLC patients, and assessment of a variety

of mutation status of multiple oncogenes has become critical in the evaluation of cancer treatments. One step RT-PCR with π Code technology enables multiplex, single-well detection of gene rearrangements from RNA specimens containing large amounts of wild-type RNA with significantly reduced sample requirement compared to conventional methods. The **IntelliPlex Lung Cancer Panel – cfRNA** identifies 28 gene variants from RNA samples (Table 1).

Table 1. Fusion Variants Detected

Gene	Fusion Variant	Inferred Breakpoint
ALK	V1	E13;A20
	V2	E20;A20
	V3a	E6a;A20
	V3b	E6b;A20
	V4	E14;A20
	V5a	E2a;A20
	V5b	E2b;A20
	V"5"	E18;A20
ROS1	CD74-ROS1	C6;R32
		C6;R34
	SLC34A2-ROS1	SL4;R32
		SL4;R34
	SDC4-ROS1	SD2;R32
		SD4;R32
SD4;R34		
EZR-ROS1	E10;R34	
TPM3-ROS1	T8;R35	
RET	KIF5B-RET	K15;R11
		K15;R12
		K16;R12
		K22;R12
		K23;R12
	NCOA4-RET	N6;R12
	TRIM33-RET	T14;R12
CCDC6-RET	C1;R12	
NTRK1	CD74-NTRK1	C8;N12
	MPRIP-NTRK1	M21;N14
MET	MET Exon14 skipping	-

3. TECHNOLOGICAL PRINCIPLES

The IntelliPlex kits utilizes two technologies, π Code and one-step RT-PCR and π Code MicroDisc, to achieve high sensitivity multiplex variants detection.

One-step RT-PCR

One-step RT-PCR combines cDNA synthesis and PCR amplification in a single tube, reducing operation time and contamination risk while yielding highly sensitive results.

π Code MicroDisc

π Code MicroDisc is manufactured to generate more than 85,000 distinct circular image patterns for multiplexing applications. Each π Code MicroDisc has a distinct circular image pattern, which corresponds to a specific capture agent conjugated to the surface of the disc. π Code tagged with different capture agents are pooled, enabling specific detection of multiple analytes in a one well reaction.

Detection Principle

The test is based on five processes:

1. RNA extraction from specimens
2. Variants-specific multiplex RT-PCR amplification
3. Hybridization of PCR amplicons with mutation/variants-specific probe tagged π Code in a one well reaction
4. Fluorescent labeling with streptavidin-phycoerythrin
5. Image pattern decoding and fluorescent signal detection by the PlexBio™ 100 Fluorescent Analyzer

4. WARNINGS AND PRECAUTIONS

- For research use only. Not for use in diagnostic procedures.
- This assay kit should only be used by qualified laboratory personnel.
- **Do not repeatedly freeze-thaw the RNA positive control. Use within two freeze-thaw cycles.**
- Separate, dedicated rooms and equipment for pre- and post- PCR process with unidirectional manner to avoid any contaminations are required.
- Pre-PCR process preparation should be operated in laminar flow hood to avoid contamination.
- Do not use a kit or reagent past its expiration date.

- Reagent components have been diluted optimally. Further dilution of the component reagents is not recommended.
- Specimens should be handled as infectious material. Please follow universal precautions for safe use.
- Store assay kits and reagents according to the product label and instructions.
- Do not mix reagents from different lots.
- Dispose of unused reagents, specimens and waste according to applicable central/federal, state, and local regulations.
- Wear powderless gloves and do not touch and make any markings on the bottom of the plate at any time, as fingerprints and markings may interfere with decoding and signal acquisition.
- General laboratory precautions should be taken:
 - Do not pipette by mouth.
 - Wear protective clothing (e.g., disposable powderless gloves and laboratory coats) and eye protection.
 - Do not eat, drink or smoke in the laboratory.
 - Wash hands thoroughly after handling samples and reagents.
- **Avoid RNase contamination:**
 - **Prepare an RNase-free working environment.**
 - **Wear gloves during all steps of the procedure.**
 - **Change gloves frequently.**
 - **Use sterile, disposable polypropylene tubes and filter tips.**
 - **Keep tubes closed whenever possible during the preparation.**
 - **Use RNase removing product to clean bench surfaces, pipettes and other components used in the experiment.**
- The workspace, including racks and pipettes, should be thoroughly cleaned and wiped with 0.5% sodium hypochlorite solution followed by wiping with a 70% ethanol solution. A 1:10 dilution of household bleach will produce a 0.5% sodium hypochlorite solution.
- Any serious incident that has occurred in relation to the device shall be reported to the manufacturer and the competent authority of the Member State in which the user and/or the patient is established.
- Material Safety Data Sheets (MSDS) are available upon request from PlexBio Customer Service.

5. KIT COMPONENTS

The **IntelliPlex Lung Cancer Panel – cfRNA** contains sufficient reagents for up to 24 tests. Kit components include:

1. **cLCP-RNA RT-PCR Buffer**
Ref. No.: 20493-R
Quantity & Volume: 1 vial, 360 µL/vial
Description: For RT-PCR amplification
Contents: 2X Reaction Mix, MgSO₄ and dNTPs
2. **cLCP-RNA RT-PCR Enzyme**
Ref. No.: 20492-R
Quantity & Volume: 1 vial, 24 µL/vial
Description: For RT-PCR amplification
Contents: RT/HotStar Taq MIX, Rnase Inhibitor (Ribolock)
3. **cLCP-RNA Primer Mix**
Ref. No.: 20491-R
Quantity & Volume: 1 vial, 96 µL/vial
Description: For RT-PCR amplification
Contents: <20 % Forward Primer, <10 % Reverse Primer (biotin labeled)
4. **cLCP-RNA πCode MicroDisc**
Ref. No.: 20494-R
Quantity & Volume: 1 vial, 480 µL/vial
Description: For PCR amplicon capture
Contents: Glycerol, πCode, Phosphate buffered saline, 0.1% Albumin from bovine (Biological), <0.1% EDTA, <0.1% Sodium azide
5. **cLCP-RNA POS Control**
Ref. No.: 20496-R
Quantity & Volume: 3 vials, lyophilized
Description: Assay positive control; reconstituted with 50 µL ddH₂O per vial prior to use.
Contents: Cell line RNA, 80 % RNAsable®
6. **NEG Control**
Ref. No.: 20549-R
Quantity & Volume: 1 vial, 120 µL/vial
Description: Assay negative control
Contents: ddH₂O
7. **cLCP-RNA Hy Buffer**
Ref. No.: 20495-R
Quantity & Volume: 1 bottle, 2.4 mL/bottle
Description: For hybridization
Contents: Saline-Sodium Phosphate-EDTA

8. **ddH₂O**
Ref. No.: 20548-R
Quantity & Volume: 1 vial, 1.5 mL/vial
Description: for reconstitution of cLCP-RNA POS Control
Contents: Nuclease-free water
9. **SA-PE Solution**
Ref. No.: 20302
Quantity & Volume: 1 bottle, 7 mL/bottle
Description: Streptavidin-phycoerythrin for fluorescent signal acquisition
Contents: Phosphate buffered saline, 0.5% Streptavidin-phycoerythrin, 1% Albumin from bovine (Biological), <0.1% Sodium azide
10. **10X Assay Wash Buffer**
Ref. No.: 20598-R
Quantity & Volume: 1 bottle, 50 mL/bottle
Description: For πCode washing
Contents: Phosphate buffered saline, 1% Tween-20, 0.5% Proclin 950

6. MATERIALS AND EQUIPMENT REQUIRED BUT NOT SUPPLIED

Required products for compatibility with IntelliPlex kits:

- 96-well plate (PlexBio; Cat. No. 80025 or Greiner Bio-one; Cat. No. 655101)
- IntelliPlex 1000 πCode Processor (PlexBio; Cat. No. 80033)
- PlexBio 100 Fluorescent Analyzer (PlexBio; Cat. No. 80000)
- U Tray (PlexBio; Cat. No. 80023)
- V Tray (PlexBio; Cat. No. 80024)
- DeXipher™ RUO (Required: PlexBio; Cat. No. 80050)

Required components:

- RNA Complete BCT® (Streck, Cat. No. 230460, 230461, 230462) or Vactainer® Venous Blood Collection Tube (BD, Cat. No. 367525)
- Qubit™ Fluorometer with dedicated quantitative reagents Qubit™ microRNA Assay Kit (Invitrogen; Cat. No. Q32880 or equivalent)
- Clean tubes for PCR reaction (Gunster; Cat. No. MB-P08A or equivalent)
- Dedicated micropipette
- Filter tips for micropipette
- ddH₂O for dilution of 10X Assay Wash Buffer
- RNA extraction kit (Recommended: QIAamp exoRNeasy Maxi Kit (50), Qiagen Cat. No. 77164 or equivalent)
- Vortex mixer

- Micro-centrifuge
- Thermocycler (Recommended: MiniAmp™ Thermal Cycler, Applied Biosystems™; Cat. No. A37834 or equivalent)
- Computer (Recommended: PlexBio; Cat. No. 80002)

7. STORAGE, STABILITY AND TRANSPORTATION

Storage

The cLCP-RNA RT-PCR Buffer and cLCP-RNA RT-PCR Enzyme of the **IntelliPlex Lung Cancer Panel – cfRNA** should be stored at -15°C to -25°C separately upon arrival.

Other kit components of the **IntelliPlex Lung Cancer Panel – cfRNA** should be stored at 2°C to 8°C.

Any reconstituted RNA positive control could be stored at -20°C up to one month. Please use within two freeze-thaw cycles.

Stability

Do not use any kit that has expired. All components are stable up to the expiration date on the label if handled and stored under the recommended conditions.

Transportation

The shipping temperature for the kit is 2-8°C. If the kit package or components is incomplete, please contact PlexBio customer service (service@plexbio.com).

8. INSTRUMENTS AND SOFTWARE

Instruments

Refer to the instrument user manuals for complete installation and operation instructions (Thermocycler, IntelliPlex 1000 π Code Processor and PlexBio 100 Fluorescent Analyzer).

Software Installation

The **IntelliPlex Lung Cancer Panel – cfRNA** has a designated Kit App and ENC file. The Kit App contains the π Code target assignments and the ENC file includes the lot number and expiration date. Please make sure you have the Kit App installed and the ENC file imported into DeXipher before your first assay run.

Kit App Installation

1. Log into www.plexbio.com and download the **LCP-cfDNA-cfRNA App**.
2. Click on the “Installer” in the APP folder and follow the instructions to complete Kit App installation.

NOTE:

The Kit App only needs to be installed once. Version

updates will be notified by customer service.

ENC File Installation

1. Log into www.plexbio.com and download the **LCP-cfDNA-cfRNA** ENC file. Each kit lot number will have a unique ENC file, so you will need to download a new ENC file each time you purchase a kit with a different lot number. Make sure to select the ENC file with the lot number that corresponds to your RUO kit.
2. Save the ENC file to your computer.
3. Follow the PlexBio 100 Fluorescent Analyzer User Manual to import the ENC file.

9. SAMPLES

Sample Collection

The **IntelliPlex Lung Cancer Panel – cfRNA** has been validated to be used for plasma. It is recommended to collect whole blood sample with Streck RNA Complete BCT® (Cat. No. 230460, 230461, 230462) or BD Vacutainer® Venous Blood Collection Tube (Cat. No. 367525). Heparin is not recommended as it may interfere with RT-PCR. The nucleic acid of all samples must be extracted with a purification procedure supporting downstream RT-PCR amplification.

Whole Blood Transportation and Storage

Blood samples in Streck RNA Complete BCT® can be transported at 15-30°C and stored within 2 days at 15-30°C according to the product specification.

Blood samples in BD Vacutainer® Venous Blood Collection Tubes can be transported at 15-25°C or 4°C and process within 1 hour.

Plasma Isolation for cfRNA extraction

To separate plasma from whole blood (8~9 mL), centrifuge the Streck RNA Complete BCT® or BD Vacutainer® Venous Blood Collection Tubes at 1,600 x g for 10 minutes at room temperature.

Transfer upper plasma layer (~4 mL) to a conical tube (not provided). Filtrate the plasma with a 0.8 μ m filter (not provided) to a new conical tube, then store in -80 °C for up to 6 months. Plasma should be transported with dry ice that keep plasma freeze during the transportation. Once arrived, move plasma into -80°C immediately.

Storage of Extracted RNA

Extracted RNA can be stored at -20°C for immediate use (≤ 24 hours), or at -80°C for long-term (1 – 14 days) storage. Do not subject the extracted RNA to repeated freeze/thaw cycles.

10. BEFORE YOU START

1. Check that the Kit App has been installed and the lot specific ENC file has been imported to DeXipher.
2. Check that you have 5 μL of extracted RNA (≥ 1 ng/ μL) ready for analysis.

11. ASSAY PROCEDURE

11.1 RNA Quantification

1. Quantify the extracted RNA using a Qubit Fluorometer with dedicated quantitative reagents (or equivalent) according to the manufacturer's protocol.
2. The RNA Stock concentration should be ≥ 1 ng/ μL to ensure optimal performance. Each RT-PCR reaction uses 10 μL of a 1 ng/ μL RNA Stock (10 ng total RNA input). RNA input amounts lower than 10 ng per reaction well are not recommended.

11.2 Reconstitute cLCP-RNA POS Control:

1. Briefly centrifuge tube.
2. Add 50 μL of ddH₂O to each required vial of cLCP-RNA POS Control.
3. Make sure the material is fully reconstituted by pipetting up and down several times.

NOTE: Leftover reconstituted POS Control can be stored at -20°C for one freeze-thaw cycle. Do not repeatedly freeze-thaw the reconstituted POS control.

11.3 Multiplex one-step RT-PCR Amplification

1. Vortex to mix each sample before use.
2. Spin down and keep samples on ice.
3. Prepare the one step RT-PCR Reaction:

For each RT-PCR reaction:

cLCP-RNA RT-PCR buffer	15 μL
cLCP-RNA RT-PCR Enzyme	1 μL
cLCP-RNA Primer Mix	4 μL
Sample/cLCP-RNA POS Control/ NEG Control	10 μL
Total	30 μL

NOTE:

- The amount of one-step RT-PCR reagent required for a Master Mix depends on the number of reactions. Always prepare a surplus.
 - Both POS Control and NEG Control are required for test validity and report generation and must be included in each assay run.
4. Mix by tapping the tubes and spin down before placing the tubes on the thermocycler. Set up the one-step RT-PCR program conditions as below:

RT-PCR Program Conditions*

Temp. (°C)	Duration	Cycles
55	15 min	1
95	2 min	1
95	15 sec	45
60	30 sec	
72	30 sec	
4	Hold	1

NOTE: Ramp rate: 3.0°C/sec (ABI MiniAmp™; Cat. No. A37834).

11.4 Hybridization and SA-PE Reaction

1. **Prepare 1X Assay Wash Buffer:** Transfer 50 mL of the 10X Assay Wash Buffer to the IntelliPlex 1000 π Code Processor 1L Assay Wash Buffer bottle and add 450 ml ddH₂O to prepare the 1X Assay Wash Buffer. Mix by swirling.

NOTE: The prepared 1X Assay Wash Buffer can be used for up to one week.

IntelliPlex 1000 π Code Processor Assay Wash Buffer consumption:

Procedure	Assay Wash Buffer Consumption (mL)
Self-test	50
DNA & RNA program (1 lane, up to 8 tests)	150
DNA & RNA program (3 lanes, up to 24 tests)	220

2. **Add 20uL cLCP – RNA πCode MicroDisc to 96 well plate:** Mix the tube of πCode by vortexing for 10 seconds, then, by pipetting, add 20 μL πCode to each well directly. Vortex the tube of πCode every four wells in between dispensing to ensure homogeneous suspension.

NOTE : Each amplified PCR products (including samples, POS and NEG control) should be added into wells lane wise, in order of A1, B1...H1 and followed by A2, B2...H2 and so on.

3. **Add 100 μL of cLCP – RNA Hy Buffer** to each well.
4. Spin down the PCR products.
5. **Denature the PCR products** on the thermocycler by heating at 95°C for 7 minutes, immediately cooled on ice/cooler or thermocycler to ensure the denatured status. Spin down before use. Use immediately (within 1 hour after denaturation).

NOTE: Pay attention to the lid temperature of thermocycler while taking out the denatured PCR products.

6. **Add 10 μL of the denatured PCR products** to each well and place the 96-well plate into the IntelliPlex 1000 πCode Processor.
7. **Pipet the desired volume of SA-PE solution** into the V Tray in SA-PE tank. Please note that the dead volume of the V Tray is **500 μL** for up to 6 selected lanes or **800 μL** if more than 6 lanes are selected. The minimum usage unit of SA-PE is **one lane (900 μL)**.

Calculation Example:

For a 3-lane reaction, the required SA-PE solution volume is at least:

$$400 \mu\text{L} \times 3 \text{ lanes} + 500 \mu\text{L}(\text{dead volume}) = 1.7 \text{ mL}$$

It is recommended to add extra solution volume into the V Tray to ensure sufficient dispensing volume.

NOTE:

Required SA-PE Solution by Lane(s):

Number of Processed Lane(s)	Required SA-PE Solution (μL)
1	900
2	1300
3	1700
4	2100
5	2500
6	2900
7	3600
8	4000
9	4400
10	4800
11	5200
12	5600

- SA-PE solution should be kept in the dark.
 - **Do not** reuse the leftover SA-PE solution and V Tray tank. Replace a new V Tray with every assay run.
8. **Run hybridization and wash:** This assay uses the **DNA/RNA program** in the **Molecular Assay** window of the IntelliPlex 1000 πCode Processor. Refer to the IntelliPlex 1000 πCode Processor operation manual and follow the instructions to run the built-in assay program as described (Homepage/ Molecular Assay/ Well Selection/ DNA/RNA / Confirm procedure conditions/ Start Running). The plate will be ready for decoding once the program is finished.

NOTE:

- IntelliPlex 1000 πCode Processor must be maintained properly and regularly.
- **Do not** open the door when the instrument is in operation.
- The kit contains sufficient reagents for 6 runs of 4 samples (including POS and NEG controls) for a maximum of 24 tests. Please note that the included Assay Wash Buffer is only sufficient for up to two independent runs. Additional Assay Wash Buffer can be ordered from PlexBio (Ref. No.: 80220).

11.5 Image Decoding and Fluorescent Detection

1. Follow the PlexBio 100 Fluorescent Analyzer User Manual to set up the read.

NOTE:

- PlexBio 100 Fluorescent Analyzer must be calibrated regularly (once per month).
 - Check that the correct ENC file has been imported.
2. Launch DeXipher to run the **Qualitative Assay**.
 3. Mark the wells for sample, positive and negative controls.
 4. Enter Assay name and place the plate into the device with the correct orientation as shown on the screen.
 5. The raw data will be analyzed through the kit ENC to generate the variant call report.

NOTE:

A single run can include from 2 to 96 tests (including POS and NEG controls) per 96 well Microwell plate. When running more than 24 specimens, multiple IntelliPlex Lung Cancer Panel – cfRNA of the same lot will be required.

12. DISCLAIMERS

Negative Test Result

A negative test result means that the targeted mutation was not detected by the kit. It does not preclude a positive result of the targeted mutation. Experimental errors or other causes may lead to false negative results. Interpretation of the results should consider these possibilities.

Positive Test Result

A positive test result means that the targeted mutation was detected by the kit. It does not preclude a negative result for the targeted mutation. Experimental errors or other causes may lead to false positive results. Interpretation of the results should consider these possibilities.

13. INTERPRETATION OF RESULT

The report generated by DeXipher includes the results of controls and samples tested in the same run. The result of external controls (POS Control and NEG Control) must be “Pass”. Otherwise, failed POS or NEG Control renders the whole assay run invalid, and the result of tested samples will not be reported.

If the result POS Control and NEG Control are “Pass”, the result of each tested sample will be reported on separate

sheets in detail. For each tested sample, its internal controls (Reference Gene Control, Internal Control, SA-PE Monitor Control, πCode MicroDiscs Count, πCode MicroDiscs Combination and Blank Control) must be “Pass”, or the test of that sample is invalid. The detection result of invalid samples will not be shown. However, failed Reference Gene Control and Internal Control do not negate samples with mutation detected. A positive sample with failed Reference Gene Control is considered valid. The detection result of target genes will be shown for each valid sample.

Refer to the chapter “**Troubleshooting**” for control failure issues.

Table 2. Interpretation of Results

Test Result	Explanation	Action
Variant Detected	Refer to Table 1	Targeted variant detected
Variant Not Detected	None	Targeted variant not detected
Result Not Shown	The test is INVALID because external controls failed OR at least one of internal controls failed	See the chapter “ Troubleshooting ” for instructions and retest.

NOTE:

“Variant Detected” indicates that the signal for at least one variant site is greater than the cutoff value of the corresponding target.

14. TROUBLESHOOTING

The troubleshooting listed below addresses possible problem causes and solutions provided during assay procedures.











Problem	Possible Cause	Recommendations
No Valid Assay Assigned	1. No plate inserted.	1. Confirm plate is inserted and repeat reading.
	2. Plate inserted in wrong orientation.	2. Confirm orientation of plate and repeat reading.
	3. No assay APP installed.	3. Install assay APP and repeat reading.
	4. No ENC file imported.	4. Import ENC file and repeat reading.
	5. Two or more lots of reagent used.	5. One reagent lot used at a time.
Positive Control Fail / Negative Control Fail	1. No POS Control or NEG Control added.	1. Ensure POS Control and NEG Control are added.
	2. RNase contamination.	2. Ensure all operating procedures are followed correctly. Ensure work environment is free of RNase.
	3. Assay did not work.	3. Make sure all the assay procedures are followed correctly.
	4. Cross contamination between samples.	4. Clean all surfaces and equipment. Operate pre-PCR and post-PCR in the dedicated area and separate the equipment for use.
	5. Wrong PC/NC wells chose.	5. Choose the correct PC/NC wells and repeat reading.

Problem	Possible Cause	Recommendations
πCode MicroDiscs Count Fail	DeXipher is unable to detect sufficient πCode MicroDiscs numbers for fusion variants determination.	
	1. πCode MicroDiscs are not proper dispersed in the well.	1. Re-disperse the microplate using IntelliPlex 1000 Processor, and repeat reading.
	2. Not enough πCode MicroDiscs added to well.	2. Ensure πCode MicroDiscs are well-mixed with proper amount added.
	3. Microbes exist in buffers.	3. Use freshly prepared assay wash buffer and ddH ₂ O for hybridization to reduce πCode MicroDiscs loss rate.
	4. Instruments error or malfunction.	4. Contact PlexBio Customer Service.
SA-PE Monitor Control Fail	Performance of SA-PE is assessed by the SAPE Monitor Control.	
	1. No SA-PE was added or insufficient SA-PE solution for dispensing.	1. Make sure all the assay procedures are followed correctly. Calculate sufficient SA-PE solution volume for dispensing. Repeat test.
	2. SA-PE solution inactivation.	2. Ensure correct storage condition and minimize the light exposure. Do not use SA-PE past its expiration date.
	3. Incorrect tested lanes of microplate selected for SA-PE solution dispensing.	3. Repeat assay and make sure lanes selected correctly.

Problem	Possible Cause	Recommendations
Blank Control Fail	"Background" is determined by measuring MFI of an internal control that should not give a signal.	
	<ol style="list-style-type: none"> Wrong hybridization conditions. Residues of SA-PE solution in wells after hybridization. PlexBio 100 Fluorescent Analyzer is not calibrated. Markings on plates. 	<ol style="list-style-type: none"> Check correct hybridization program is selected. Ensure all buffers (Assay Wash Buffer and ddH₂O) on IntelliPlex 1000 Processor are fresh-made and sufficient for washing procedures. Perform calibration on PlexBio 100 Fluorescent Analyzer. Do not make any marking on plate.
Internal Control Fail	Internal Control monitors all steps in the procedure and must be positive.	
	1. PCR inhibition exists.	1. Follow instructions of sample extraction carefully. Ensure required temperature ranges and centrifugation needs are complied. Ensure complete removal of ethanol.
	2. PCR procedures are not performed correctly.	2. Make sure all PCR procedures are followed correctly. Do not to use expired materials or mixed lots of reagents. Ensure storage conditions are correct.
	3. RNase contamination.	3. Ensure all the operating procedures are followed correctly. Ensure work environment is free of RNase.
4. Hybridization did not work.	4. Make sure all the assay procedures are followed correctly. Ensure samples are freshly heat-denatured.	

Problem	Possible Cause	Recommendations
Reference Gene Fail	Reference Gene monitors quality of tested sample and must be positive.	
	1. No Sample added or absence of human-derived RNA.	1. Ensure human-derived RNA samples are added. Do not use artificial RNA as samples which may generate invalid results.
	2. Insufficient sample input for assays or poor sample quality.	2. Quantify samples and check the sample input and RIN (RNA integrity number) value. If still remains failed, ensure the collected samples meet specimen requirements. Retest with new samples if needed.
	3. PCR inhibition exists.	3. Follow sample extraction instructions carefully. Ensure required temperature ranges and centrifugation needs are complied. Ensure complete removal of ethanol.
4. PCR procedures are not performed correctly.	4. Make sure all PCR procedures are followed correctly. Do not to use expired materials or mixed lots of reagents. Ensure storage conditions are correct.	

15. SYMBOLS

Symbol	Explanation	Symbol	Explanation
	For research use only		Catalog number
	Batch number		Consult instructions for use
	Manufacturer		Use by Date
	Temperature limitation		Caution
	Contains sufficient for <n> tests		Date of Manufacture

16. REFERENCES

1. <https://www.mycancergenome.org/content/diseases/lung-cancer/>
2. Pao W, Miller V, Zakowski M, Doherty J, Politi K, Sarkaria I, Singh B, Heelan R, Rusch V, Fulton L, Mardis E, Kupfer D, Wilson R, Kris M, Varmus H. (2004) EGFR receptor gene mutations are common in lung cancers from “never smokers” and are associated with sensitivity of tumors to gefitinib and erlotinib. *Proc Natl Acad Sci USA*. 101(36):13306-11
3. Lie`vre A, Bachet JB, Le Corre D, Boige V, Landi B, Emile JF, Co`te JF, Tomasic G, Penna C, Ducreux M, Rougier P, Penault-Llorca F, Laurent-Puig P. (2006) KRAS mutation status is predictive of response to cetuimab therapy in colorectal cancer. *Cancer Res*. 66(8):3992-5
4. <http://www.uniprot.org/uniprot/P01116>
5. Baynes RD, Gansert J. (2009) KRAS mutational status as a predictor of epidermal growth factor receptor inhibitor efficacy in colorectal cancer. *Am J Ther*. 16(6):554-61
6. Sakuma, M. (2000) PriProbit, ver. 1.63. Available from James E. Throne USDA-ARS GMPRC, Manhattan, KS (<http://bru.usgmrl.ksu.edu/throne/>)
7. Neumann J, Zeindl-Eberhart E, Kirchner T, Jung A. (2009) Frequency and type of KRAS mutations in routine diagnostic analysis of metastatic colorectal cancer. *Pathol Res Pract*. 2009; 205:858–862.
8. Food and Drug Administration. Class labeling changes to anti-EGFR monoclonal antibodies, cetuximab (Erbix) and panitumumab (Vectibix): KRAS mutations.


<http://www.fda.gov/AboutFDA/CentersOffices/CDER/ucml172905.htm>.

9. Loupakis F, Ruzzo A, Cremolini C, et al. KRAS codon 61, 146 and BRAF mutations predict resistance to cetuximab plus irinotecan in KRAS codon 12 and 13 wild-type metastatic colorectal cancer. *Br J Cancer* 2009 Aug 18;101(4):715-21.
10. Siena S, Sartore-Bianchi A, Di Nicolantonio F, et al. Biomarkers predicting clinical outcome of epidermal growth factor receptortargeted therapy in metastatic colorectal cancer. *J Natl Cancer Inst* 2009 Oct 7;101(19):1308-24.

Notice to User

The use of this product and the associated PlexBio instrumentation is covered by one or more issued (US10302640B2, US10436778B2, US10436776B2, US9063044B2, US10019815B2) and pending US and foreign patents owned by PlexBio Co., Ltd. The purchase of this product includes nontransferable rights to use this amount of the product to practice the methods described therein. No general patent or other license of any kind other than this specific right of use from purchase is granted. Further information on purchasing licenses for other applications can be obtained from PlexBio Co., Ltd. 6F-1, No. 351, Yangguang St., Neihu District, Taipei City 11491, Taiwan.

Trademarks

 , π Code™ , DeXipher™, IntelliPlex™, DigiPlex™, π Code™ , PlexBio™, are trademarks or registered trademarks of PlexBio Co., Ltd. All other product names and any registered and unregistered trademarks mentioned are used for identification purposes only and remain the exclusive property of their respective owners.

Copyright

©2020 PlexBio Co., Ltd. All rights reserved.